



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/638,426	08/12/2003	Samuel J. Epstein	12013/47601	5077
23838	7590	06/13/2007		
KENYON & KENYON LLP 1500 K STREET N.W. SUITE 700 WASHINGTON, DC 20005			EXAMINER KOHARSKI, CHRISTOPHER	
			ART UNIT 3763	PAPER NUMBER
			MAIL DATE 06/13/2007	DELIVERY MODE PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary

Application No.

10/638,426

Applicant(s)

EPSTEIN ET AL.

Examiner

Christopher D. Koharski

Art Unit

3763

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 11/20/2006.
- 2a) ☒ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-21 is/are pending in the application.
- 4a) Of the above claim(s) 3,4,6-18 and 20 is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1,2,5,19 and 21 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
 - ☐ Certified copies of the priority documents have been received in Application No. _____.
 - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- 1) ☒ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☐ Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)
Paper No(s)/Mail Date _____
- 4) ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____
- 5) ☐ Notice of Informal Patent Application (PTO-152)
- 6) ☐ Other: _____

DETAILED ACTION

Response to Amendment

Examiner acknowledges the reply filed 11/20/2007 in which no claims were amended, currently claims 1-21 are pending for examination, with claims 3, 4, 6-18 and 20 are withdrawn from a previous election restriction.

Claim Rejections - 35 USC § 102

1. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

2. Claims 1, 2, 5, 19, and 21 are rejected under 35 U.S.C. 102(b) as being anticipated by Nilsson et al (US 6,132,405). Nilsson et al. discloses a catheter for peritoneal dialysis.

Regarding claims 1, 2, 5, 19, and 21, Nilsson et al. discloses a device (figs. 3, 6, and 7) and method for delivery of non-Newtonian fluid (col 1, ln 5-10; col 2, ln 25-40; col 4, ln 30-35) for therapeutic purposes.

Response to Arguments

Applicant's arguments filed 11/20/2006 have been fully considered but they are not persuasive. Applicant's Representative asserts that the Nilsson et al (US 6,132,405) reference does not disclose a device and method for delivering a non-Newtonian fluid.

Examiner has fully considered applicant's arguments but they are not persuasive. It is examiners position that given a careful reading, the claims do not distinguish over the prior art of record.

Examiner disagrees, specifically regarding claim 1, the catheter of Nilsson et al. is capable of delivering a non-Newtonian fluid, it is well established that a recitation with respect to the manner in which an apparatus is intended to be employed, i.e., a functional limitation, does not impose any structural limitation upon the claimed apparatus which differentiates it from a prior art reference disclosing the structural limitations of the claim, see *In re Pearson*, 494 F.2d 1399, 181 USPQ 641 (CCPA 1974). Specifically regarding claim 19, Nilsson et al. discloses a catheter and method that is used with non-Newtonian fluids (see above rejection) such as blood, urine or dialysis fluid. The supported definition of a non-Newtonian fluid is: "a fluid in which the viscosity changes with the applied strain rate. As a result, non-Newtonian fluids may not have a well-defined viscosity" given examples are blood, milk, concentrated dissolved sugars/salts in water, colloids, clay, gelatin, etc (wikipedia.com).

Therefore, the prior art of record teaches all elements as claimed and these elements satisfy all structural, functional, operational, and spatial limitations currently in the claims. Therefore the standing rejections are proper and maintained.

Conclusion

3. The prior art made of record and not relied upon is considered pertinent to applicant's disclosure. See PTO-892 Form.

THIS ACTION IS MADE FINAL. Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire **THREE MONTHS** from the mailing date of this action. In the event a first reply is filed within **TWO MONTHS** of the mailing date of this final action and the advisory action is not mailed until after the end of the **THREE-MONTH** shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than **SIX MONTHS** from the mailing date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Christopher D. Koharski whose telephone number is 571-272-7230. The examiner can normally be reached on 7:30am to 4:00pm EST.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Nick Lucchesi can be reached on 571-272-4977. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3763

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

Date:

6/9/07



Christopher D. Koharski
AU 3763



NICHOLAS D. LUCCHESI
SUPERVISORY PATENT EXAMINER
TECHNOLOGY CENTER 3700